



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18134

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0000	1.0000
10.00	1.0000	1.0000	1.0000
20.00	0.9995	1.0000	1.0000
30.00	0.9987	1.0001	1.0000
40.00	0.9977	1.0002	1.0001
50.00	0.9966	1.0005	1.0003
60.00	0.9954	1.0009	1.0006
70.00	0.9946	1.0016	1.0011
80.00	0.9940	1.0028	1.0020
90.00	0.9942	1.0045	1.0034
100.00	0.9952	1.0071	1.0054
120.00	1.0010	1.0156	1.0120
140.00	1.0132	1.0301	1.0228
160.00	1.0275	1.0460	1.0333

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.501	2.499
Cross-axis Coupling (g/g)	0.20%	0.12%	0.20%