The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 18135

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0000	1.0001	0.9998
10.00	1.0000	1.0000	1.0000
20.00	1.0001	0.9998	1.0006
30.00	1.0003	0.9995	1.0016
40.00	1.0005	0.9992	1.0026
50.00	1.0007	0.9990	1.0037
60.00	1.0009	0.9989	1.0045
70.00	1.0011	0.9991	1.0051
80.00	1.0015	0.9998	1.0054
90.00	1.0021	1.0012	1.0054
100.00	1.0033	1.0033	1.0054
120.00	1.0087	1.0112	1.0069
140.00	1.0209	1.0250	1.0142
160.00	1.0376	1.0403	1.0287

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.500
Cross-axis Coupling (g/g)	0.04%	0.28%	0.16%