The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 18142

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0003	1.0004
10.00	1.0000	1.0000	1.0000
20.00	0.9995	0.9990	0.9988
30.00	0.9988	0.9975	0.9970
40.00	0.9980	0.9957	0.9951
50.00	0.9974	0.9939	0.9934
60.00	0.9970	0.9924	0.9924
70.00	0.9972	0.9915	0.9925
80.00	0.9981	0.9915	0.9940
90.00	1.0001	0.9928	0.9971
100.00	1.0031	0.9955	1.0017
120.00	1.0139	1.0065	1.0153
140.00	1.0319	1.0258	1.0310
160.00	1.0529	1.0475	1.0394

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.501	2.499
Cross-axis Coupling (g/g)	0.44%	0.52%	0.20%