



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18145

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9999	1.0003	0.9999
10.00	1.0000	1.0000	1.0000
20.00	1.0004	0.9992	1.0002
30.00	1.0010	0.9980	1.0005
40.00	1.0016	0.9966	1.0008
50.00	1.0020	0.9953	1.0013
60.00	1.0022	0.9945	1.0018
70.00	1.0019	0.9943	1.0025
80.00	1.0012	0.9952	1.0035
90.00	1.0001	0.9973	1.0047
100.00	0.9990	1.0006	1.0065
120.00	0.9975	1.0119	1.0127
140.00	1.0009	1.0276	1.0238
160.00	1.0110	1.0394	1.0372

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.502	2.500
Cross-axis Coupling (g/g)	0.24%	0.32%	0.04%