The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 18148

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9994	1.0003	1.0003
10.00	1.0000	1.0000	1.0000
20.00	1.0016	0.9991	0.9992
30.00	1.0037	0.9976	0.9978
40.00	1.0058	0.9958	0.9961
50.00	1.0072	0.9940	0.9943
60.00	1.0079	0.9923	0.9924
70.00	1.0078	0.9911	0.9909
80.00	1.0071	0.9906	0.9899
90.00	1.0065	0.9914	0.9898
100.00	1.0065	0.9937	0.9910
120.00	1.0107	1.0043	0.9984
140.00	1.0236	1.0252	1.0152
160.00	1.0408	1.0506	1.0376

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.501	2.500
Cross-axis Coupling (g/g)	0.28%	0.20%	0.24%