



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18149

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9999	1.0001	1.0002
10.00	1.0000	1.0000	1.0000
20.00	1.0003	0.9997	0.9996
30.00	1.0008	0.9992	0.9990
40.00	1.0014	0.9987	0.9985
50.00	1.0020	0.9984	0.9982
60.00	1.0026	0.9982	0.9985
70.00	1.0030	0.9986	0.9995
80.00	1.0033	0.9996	1.0017
90.00	1.0035	1.0013	1.0049
100.00	1.0038	1.0041	1.0095
120.00	1.0049	1.0127	1.0223
140.00	1.0083	1.0244	1.0365
160.00	1.0136	1.0309	1.0400

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.500
Cross-axis Coupling (g/g)	0.36%	0.60%	0.48%