



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14006

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0020	1.0019
10.00	1.0000	1.0000	1.0000
20.00	0.9950	0.9945	0.9947
30.00	0.9880	0.9867	0.9873
40.00	0.9805	0.9783	0.9792
50.00	0.9745	0.9714	0.9724
60.00	0.9713	0.9674	0.9685
70.00	0.9728	0.9681	0.9692
80.00	0.9806	0.9753	0.9763
90.00	0.9949	0.9892	0.9901
100.00	1.0166	1.0107	1.0117
120.00	1.0810	1.0756	1.0790
140.00	1.1406	1.1403	1.1519
160.00	1.1003	1.1134	1.1315

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.499	2.502	2.497
Cross-axis Coupling (g/g)	0.20%	0.76%	0.24%