



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14007

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0019	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9939	0.9948	0.9950
30.00	0.9854	0.9873	0.9878
40.00	0.9760	0.9791	0.9799
50.00	0.9681	0.9721	0.9730
60.00	0.9632	0.9677	0.9686
70.00	0.9633	0.9678	0.9681
80.00	0.9700	0.9739	0.9733
90.00	0.9839	0.9866	0.9846
100.00	1.0059	1.0069	1.0029
120.00	1.0744	1.0722	1.0620
140.00	1.1467	1.1484	1.1310
160.00	1.1227	1.1410	1.1250

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.507	2.499	2.499
Cross-axis Coupling (g/g)	0.36%	0.24%	0.68%