



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14011

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0035	1.0031	1.0031
10.00	1.0000	1.0000	1.0000
20.00	0.9903	0.9915	0.9916
30.00	0.9773	0.9802	0.9804
40.00	0.9641	0.9688	0.9692
50.00	0.9540	0.9601	0.9609
60.00	0.9485	0.9554	0.9568
70.00	0.9494	0.9559	0.9583
80.00	0.9576	0.9625	0.9665
90.00	0.9725	0.9745	0.9807
100.00	0.9941	0.9919	1.0012
120.00	1.0550	1.0409	1.0609
140.00	1.1157	1.0914	1.1271
160.00	1.1022	1.0898	1.1276

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.500	2.500	2.499
Cross-axis Coupling (g/g)	0.16%	0.16%	0.24%