



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14013

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0036	1.0033	1.0031
10.00	1.0000	1.0000	1.0000
20.00	0.9904	0.9909	0.9917
30.00	0.9785	0.9785	0.9805
40.00	0.9675	0.9658	0.9693
50.00	0.9604	0.9557	0.9609
60.00	0.9582	0.9502	0.9567
70.00	0.9616	0.9508	0.9580
80.00	0.9716	0.9588	0.9657
90.00	0.9869	0.9737	0.9791
100.00	1.0080	0.9958	0.9981
120.00	1.0654	1.0585	1.0508
140.00	1.1225	1.1176	1.1017
160.00	1.1104	1.0913	1.0885

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.496	2.497	2.504
Cross-axis Coupling (g/g)	0.36%	0.12%	0.28%