



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14016

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0040	1.0035	1.0042
10.00	1.0000	1.0000	1.0000
20.00	0.9897	0.9908	0.9892
30.00	0.9771	0.9791	0.9758
40.00	0.9655	0.9681	0.9632
50.00	0.9577	0.9603	0.9541
60.00	0.9543	0.9568	0.9491
70.00	0.9560	0.9582	0.9491
80.00	0.9632	0.9652	0.9549
90.00	0.9752	0.9772	0.9663
100.00	0.9923	0.9941	0.9839
120.00	1.0412	1.0414	1.0388
140.00	1.0934	1.0873	1.1028
160.00	1.0882	1.0722	1.0968

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.499	2.498
Cross-axis Coupling (g/g)	0.60%	0.16%	0.20%