



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14017

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0021	1.0019
10.00	1.0000	1.0000	1.0000
20.00	0.9904	0.9914	0.9918
30.00	0.9789	0.9732	0.9768
40.00	0.9666	0.9687	0.9640
50.00	0.9545	0.9535	0.9563
60.00	0.9524	0.9512	0.9534
70.00	0.9557	0.9515	0.9523
80.00	0.9623	0.9604	0.9580
90.00	0.9747	0.9711	0.9710
100.00	0.9912	0.9914	0.9945
120.00	1.0415	1.0526	1.0440
140.00	1.0924	1.1154	1.1025
160.00	1.0875	1.1020	1.0945

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.497	2.501	2.498
Cross-axis Coupling (g/g)	0.16%	0.16%	0.04%