



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14035

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0017	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9946	0.9953	0.9948
30.00	0.9868	0.9884	0.9872
40.00	0.9782	0.9805	0.9786
50.00	0.9705	0.9731	0.9706
60.00	0.9653	0.9672	0.9646
70.00	0.9643	0.9646	0.9621
80.00	0.9693	0.9662	0.9644
90.00	0.9808	0.9727	0.9722
100.00	1.0000	0.9848	0.9862
120.00	1.0642	1.0262	1.0335
140.00	1.1436	1.0756	1.0907
160.00	1.1414	1.0725	1.0902

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.497	2.501	2.497
Cross-axis Coupling (g/g)	0.44%	0.16%	0.28%