



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14036

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0024	1.0015	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9932	0.9957	0.9953
30.00	0.9835	0.9897	0.9886
40.00	0.9727	0.9832	0.9810
50.00	0.9631	0.9782	0.9745
60.00	0.9564	0.9759	0.9701
70.00	0.9545	0.9781	0.9696
80.00	0.9590	0.9866	0.9745
90.00	0.9705	1.0016	0.9855
100.00	0.9897	1.0244	1.0036
120.00	1.0501	1.0934	1.0639
140.00	1.1084	1.1621	1.1406
160.00	1.0715	1.1228	1.1494

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.498	2.498	2.498
Cross-axis Coupling (g/g)	0.24%	0.36%	0.12%