The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/INT P/N:97837
S/N: 14044

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0024	1.0026	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9934	0.9926	0.9941
30.00	0.9839	0.9821	0.9857
40.00	0.9734	0.9703	0.9763
50.00	0.9642	0.9600	0.9680
60.00	0.9577	0.9525	0.9623
70.00	0.9562	0.9503	0.9612
80.00	0.9612	0.9549	0.9663
90.00	0.9735	0.9669	0.9782
100.00	0.9939	0.9872	0.9980
120.00	1.0604	1.0522	1.0629
140.00	1.1340	1.1156	1.1373
160.00	1.1132	1.0725	1.1210

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.501	2.500	2.501
Cross-axis Coupling (g/g)	0.12%	0.16%	0.12%