The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/INT P/N:97837
S/N: 14047

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0022	1.0023
10.00	1.0000	1.0000	1.0000
20.00	0.9954	0.9938	0.9935
30.00	0.9887	0.9850	0.9842
40.00	0.9812	0.9750	0.9737
50.00	0.9744	0.9661	0.9641
60.00	0.9696	0.9597	0.9570
70.00	0.9683	0.9577	0.9543
80.00	0.9717	0.9618	0.9576
90.00	0.9807	0.9726	0.9675
100.00	0.9959	0.9912	0.9851
120.00	1.0460	1.0531	1.0443
140.00	1.1044	1.1255	1.1119
160.00	1.0969	1.1126	1.0938

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.499	2.502	2.500
Cross-axis Coupling (g/g)	0.04%	0.00%	0.00%