



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14048

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0015	1.0019	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9958	0.9948	0.9943
30.00	0.9898	0.9872	0.9861
40.00	0.9832	0.9788	0.9766
50.00	0.9775	0.9714	0.9679
60.00	0.9740	0.9662	0.9610
70.00	0.9739	0.9650	0.9579
80.00	0.9785	0.9690	0.9595
90.00	0.9883	0.9789	0.9667
100.00	1.0035	0.9952	0.9800
120.00	1.0501	1.0472	1.0260
140.00	1.0977	1.1036	1.0819
160.00	1.0832	1.0896	1.0824

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.503	2.500	2.497
Cross-axis Coupling (g/g)	0.60%	0.32%	0.24%