



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 14051

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0012	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9939	0.9965	0.9954
30.00	0.9852	0.9915	0.9888
40.00	0.9754	0.9859	0.9813
50.00	0.9666	0.9810	0.9744
60.00	0.9602	0.9778	0.9693
70.00	0.9582	0.9778	0.9675
80.00	0.9619	0.9820	0.9703
90.00	0.9719	0.9912	0.9784
100.00	0.9890	1.0061	0.9923
120.00	1.0445	1.0550	1.0390
140.00	1.1037	1.1161	1.0946
160.00	1.0838	1.1234	1.0917

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.501	2.500	2.502
Cross-axis Coupling (g/g)	0.04%	0.40%	0.04%