The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/INT P/N:97837
S/N: 15001

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0018	1.0024
10.00	1.0000	1.0000	1.0000
20.00	0.9951	0.9949	0.9933
30.00	0.9882	0.9876	0.9838
40.00	0.9805	0.9794	0.9733
50.00	0.9739	0.9722	0.9642
60.00	0.9697	0.9672	0.9580
70.00	0.9695	0.9662	0.9567
80.00	0.9748	0.9706	0.9619
90.00	0.9860	0.9810	0.9739
100.00	1.0039	0.9984	0.9935
120.00	1.0600	1.0556	1.0548
140.00	1.1213	1.1268	1.1168
160.00	1.1095	1.1327	1.0924

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.498	2.504
Cross-axis Coupling (g/g)	0.16%	0.32%	0.80%