



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 15030

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0021	1.0022	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9942	0.9938	0.9941
30.00	0.9857	0.9847	0.9857
40.00	0.9760	0.9744	0.9762
50.00	0.9670	0.9649	0.9677
60.00	0.9600	0.9576	0.9617
70.00	0.9570	0.9545	0.9601
80.00	0.9591	0.9568	0.9642
90.00	0.9672	0.9656	0.9749
100.00	0.9822	0.9819	0.9931
120.00	1.0351	1.0395	1.0544
140.00	1.1038	1.1168	1.1324
160.00	1.1096	1.1288	1.1409

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.498	2.498
Cross-axis Coupling (g/g)	0.80%	0.56%	0.88%