



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 15035

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0023	1.0021	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9935	0.9940	0.9945
30.00	0.9841	0.9855	0.9865
40.00	0.9736	0.9761	0.9776
50.00	0.9643	0.9679	0.9696
60.00	0.9577	0.9624	0.9640
70.00	0.9558	0.9618	0.9628
80.00	0.9599	0.9677	0.9673
90.00	0.9708	0.9809	0.9786
100.00	0.9892	1.0027	0.9977
120.00	1.0483	1.0749	1.0622
140.00	1.1110	1.1597	1.1420
160.00	1.0883	1.1388	1.1358

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.500
Cross-axis Coupling (g/g)	0.20%	0.04%	0.84%