



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 15040

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0017	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9955	0.9953	0.9952
30.00	0.9890	0.9886	0.9883
40.00	0.9817	0.9812	0.9807
50.00	0.9752	0.9749	0.9741
60.00	0.9707	0.9710	0.9698
70.00	0.9699	0.9714	0.9695
80.00	0.9742	0.9777	0.9750
90.00	0.9843	0.9904	0.9867
100.00	1.0011	1.0108	1.0060
120.00	1.0578	1.0770	1.0696
140.00	1.1292	1.1557	1.1479
160.00	1.1320	1.1458	1.1449

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.499
Cross-axis Coupling (g/g)	0.08%	0.44%	0.36%