The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/INT P/N:97837
S/N: 15046

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0025	1.0023
10.00	1.0000	1.0000	1.0000
20.00	0.9937	0.9931	0.9935
30.00	0.9846	0.9832	0.9842
40.00	0.9744	0.9722	0.9738
50.00	0.9653	0.9624	0.9646
60.00	0.9587	0.9555	0.9583
70.00	0.9568	0.9535	0.9567
80.00	0.9609	0.9580	0.9617
90.00	0.9720	0.9699	0.9738
100.00	0.9910	0.9902	0.9941
120.00	1.0547	1.0581	1.0611
140.00	1.1311	1.1406	1.1380
160.00	1.1206	1.1349	1.1182

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.500
Cross-axis Coupling (g/g)	0.20%	0.08%	0.04%