



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N: 97837

S/N: 15058

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0013	1.0021	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9962	0.9942	0.9952
30.00	0.9908	0.9858	0.9882
40.00	0.9849	0.9765	0.9805
50.00	0.9800	0.9684	0.9738
60.00	0.9771	0.9630	0.9695
70.00	0.9779	0.9622	0.9693
80.00	0.9837	0.9676	0.9750
90.00	0.9953	0.9800	0.9872
100.00	1.0136	1.0004	1.0074
120.00	1.0731	1.0671	1.0753
140.00	1.1487	1.1455	1.1656
160.00	1.1596	1.1342	1.1798

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.503	2.502
Cross-axis Coupling (g/g)	0.28%	0.52%	0.20%