



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 15065

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0021	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9945	0.9942	0.9943
30.00	0.9866	0.9857	0.9859
40.00	0.9775	0.9761	0.9762
50.00	0.9692	0.9674	0.9673
60.00	0.9628	0.9608	0.9602
70.00	0.9603	0.9584	0.9571
80.00	0.9629	0.9613	0.9592
90.00	0.9718	0.9706	0.9678
100.00	0.9878	0.9871	0.9840
120.00	1.0449	1.0436	1.0434
140.00	1.1238	1.1130	1.1274
160.00	1.1435	1.1086	1.1423

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.501
Cross-axis Coupling (g/g)	0.56%	0.28%	0.08%