



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 15069

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0019	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9939	0.9948	0.9940
30.00	0.9852	0.9874	0.9855
40.00	0.9755	0.9793	0.9760
50.00	0.9671	0.9724	0.9677
60.00	0.9613	0.9684	0.9621
70.00	0.9601	0.9689	0.9611
80.00	0.9652	0.9757	0.9664
90.00	0.9774	0.9893	0.9788
100.00	0.9976	1.0108	0.9994
120.00	1.0648	1.0797	1.0677
140.00	1.1465	1.1589	1.1493
160.00	1.1414	1.1443	1.1349

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.503
Cross-axis Coupling (g/g)	0.60%	0.40%	0.08%