



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 15071

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0026	1.0019	1.0024
10.00	1.0000	1.0000	1.0000
20.00	0.9927	0.9948	0.9931
30.00	0.9823	0.9873	0.9833
40.00	0.9705	0.9791	0.9723
50.00	0.9599	0.9721	0.9626
60.00	0.9518	0.9676	0.9557
70.00	0.9484	0.9675	0.9537
80.00	0.9511	0.9734	0.9582
90.00	0.9608	0.9860	0.9700
100.00	0.9782	1.0063	0.9903
120.00	1.0385	1.0730	1.0585
140.00	1.1161	1.1570	1.1436
160.00	1.1297	1.1642	1.1448

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.499
Cross-axis Coupling (g/g)	0.48%	0.16%	0.20%