



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT

P/N:97837

S/N: 15072

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0025	1.0025
10.00	1.0000	1.0000	1.0000
20.00	0.9939	0.9929	0.9930
30.00	0.9851	0.9829	0.9829
40.00	0.9753	0.9720	0.9716
50.00	0.9666	0.9626	0.9614
60.00	0.9606	0.9564	0.9538
70.00	0.9592	0.9557	0.9511
80.00	0.9639	0.9621	0.9548
90.00	0.9755	0.9761	0.9658
100.00	0.9950	0.9990	0.9852
120.00	1.0600	1.0719	1.0532
140.00	1.1422	1.1532	1.1451
160.00	1.1510	1.1383	1.1637

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.502
Cross-axis Coupling (g/g)	0.24%	0.40%	0.20%