The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/1-H
S/N: 81116

Normalized Magnitude-frequency Characteristics

Freq. (Hz)	X-Axis (Horizontal)
1.00	1.0003
10.00	1.0000
20.00	0.9991
30.00	0.9977
40.00	0.9958
50.00	0.9937
60.00	0.9915
70.00	0.9896
80.00	0.9882
90.00	0.9879
100.00	0.9891
120.00	0.9983
140.00	1.0212
160.00	1.0556

Output Sensitivity

	X-Axis (Horizontal)
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s^2