



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2018/12/21

REF TEK 147A-01/1/H

S/N:81171

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0002
10.00	1.0000
20.00	0.9994
30.00	0.9984
40.00	0.9971
50.00	0.9958
60.00	0.9945
70.00	0.9935
80.00	0.9929
90.00	0.9932
100.00	0.9944
120.00	1.0010
140.00	1.0143
160.00	1.0296

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²