



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2018/12/21

REF TEK 147A-01/1/H

S/N:81178

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	0.9999
10.00	1.0000
20.00	1.0003
30.00	1.0007
40.00	1.0013
50.00	1.0018
60.00	1.0023
70.00	1.0027
80.00	1.0033
90.00	1.0040
100.00	1.0053
120.00	1.0109
140.00	1.0239
160.00	1.0428

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²