



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2018/12/21

REF TEK 147A-01/1/H

S/N:81179

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0002
10.00	1.0000
20.00	0.9995
30.00	0.9988
40.00	0.9977
50.00	0.9965
60.00	0.9952
70.00	0.9941
80.00	0.9932
90.00	0.9930
100.00	0.9939
120.00	1.0007
140.00	1.0178
160.00	1.0418

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.502

* g = 9.80 m/s²