



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/07

REF TEK 147A-01/1/H

S/N:81189

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	0.9999
10.00	1.0000
20.00	1.0003
30.00	1.0011
40.00	1.0024
50.00	1.0041
60.00	1.0062
70.00	1.0082
80.00	1.0099
90.00	1.0107
100.00	1.0107
120.00	1.0092
140.00	1.0103
160.00	1.0205

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.502

* g = 9.80 m/s²