



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/07

REF TEK 147A-01/1/H

S/N:81193

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0002
10.00	1.0000
20.00	0.9993
30.00	0.9982
40.00	0.9968
50.00	0.9951
60.00	0.9933
70.00	0.9917
80.00	0.9904
90.00	0.9899
100.00	0.9904
120.00	0.9962
140.00	1.0115
160.00	1.0334

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.502

* g = 9.80 m/s²