



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/08

REF TEK 147A-01/1/H

S/N:81204

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	0.9993
10.00	1.0000
20.00	1.0018
30.00	1.0038
40.00	1.0051
50.00	1.0051
60.00	1.0036
70.00	1.0009
80.00	0.9973
90.00	0.9937
100.00	0.9907
120.00	0.9888
140.00	0.9959
160.00	1.0114

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²