



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/09

REF TEK 147A-01/1/H

S/N:81209

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	0.9986
10.00	1.0000
20.00	1.0032
30.00	1.0065
40.00	1.0084
50.00	1.0083
60.00	1.0064
70.00	1.0033
80.00	0.9996
90.00	0.9963
100.00	0.9938
120.00	0.9937
140.00	1.0028
160.00	1.0188

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.502

* g = 9.80 m/s²