



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/10

REF TEK 147A-01/1/H
S/N:81222

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0001
10.00	1.0000
20.00	0.9998
30.00	0.9994
40.00	0.9989
50.00	0.9981
60.00	0.9969
70.00	0.9955
80.00	0.9937
90.00	0.9918
100.00	0.9900
120.00	0.9878
140.00	0.9918
160.00	1.0056

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²