



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/28

REF TEK 147A-01/1/H

S/N:81233

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	0.9990
10.00	1.0000
20.00	1.0028
30.00	1.0064
40.00	1.0096
50.00	1.0113
60.00	1.0108
70.00	1.0081
80.00	1.0035
90.00	0.9985
100.00	0.9943
120.00	0.9933
140.00	1.0101
160.00	1.0418

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²