



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/29

REF TEK 147A-01/1/H
S/N:81240

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0002
10.00	1.0000
20.00	0.9995
30.00	0.9988
40.00	0.9980
50.00	0.9970
60.00	0.9961
70.00	0.9955
80.00	0.9953
90.00	0.9957
100.00	0.9971
120.00	1.0036
140.00	1.0173
160.00	1.0361

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²