



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/01/30

REF TEK 147A-01/1/H

S/N:81242

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	0.9989
10.00	1.0000
20.00	1.0030
30.00	1.0069
40.00	1.0107
50.00	1.0131
60.00	1.0138
70.00	1.0128
80.00	1.0104
90.00	1.0077
100.00	1.0054
120.00	1.0055
140.00	1.0170
160.00	1.0378

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²