



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/02/25

REF TEK 147A-01/1/H

S/N:81269

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0001
10.00	1.0000
20.00	0.9996
30.00	0.9989
40.00	0.9979
50.00	0.9967
60.00	0.9953
70.00	0.9938
80.00	0.9922
90.00	0.9910
100.00	0.9901
120.00	0.9908
140.00	0.9973
160.00	1.0097

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²