



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/03/05

REF TEK 147A-01/1/H

S/N:81305

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0000
10.00	1.0000
20.00	0.9999
30.00	0.9997
40.00	0.9995
50.00	0.9991
60.00	0.9985
70.00	0.9980
80.00	0.9974
90.00	0.9973
100.00	0.9976
120.00	1.0018
140.00	1.0143
160.00	1.0350

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²