



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/03/05

REF TEK 147A-01/1/H

S/N:81306

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	0.9995
10.00	1.0000
20.00	1.0015
30.00	1.0034
40.00	1.0052
50.00	1.0063
60.00	1.0063
70.00	1.0050
80.00	1.0026
90.00	0.9999
100.00	0.9978
120.00	0.9991
140.00	1.0148
160.00	1.0433

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²