



# REF TEK

SYSTEMS INC.

**The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.**

**Reftek Systems Inc. purchased the REF TEK product line in May 2020.**

**Reftek Systems will honour existing product warranties and extended warranties.**

**ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:**

**Sales & Administrative:** [sales@reftek.com](mailto:sales@reftek.com)

**Support & Engineering:** [support@reftek.com](mailto:support@reftek.com)

**Accounting:** [accounting@reftek.com](mailto:accounting@reftek.com)

**Address:** 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

**Phone:** +1-902-444-7650

**Fax:** +1-902-444-7651

DATE: 2019/03/06

REF TEK 147A-01/1/H  
S/N:81320

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

| Freq. (Hz) | Horizontal |
|------------|------------|
| 1.00       | 1.0002     |
| 10.00      | 1.0000     |
| 20.00      | 0.9993     |
| 30.00      | 0.9983     |
| 40.00      | 0.9969     |
| 50.00      | 0.9954     |
| 60.00      | 0.9939     |
| 70.00      | 0.9926     |
| 80.00      | 0.9915     |
| 90.00      | 0.9911     |
| 100.00     | 0.9914     |
| 120.00     | 0.9956     |
| 140.00     | 1.0065     |
| 160.00     | 1.0230     |

OUTPUT SENSITIVITY

| Horizontal               |       |
|--------------------------|-------|
| Output Sensitivity (V/g) | 2.500 |

\* g = 9.80 m/s<sup>2</sup>