



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/03/06

REF TEK 147A-01/1/H

S/N:81323

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0000
10.00	1.0000
20.00	0.9998
30.00	0.9996
40.00	0.9991
50.00	0.9985
60.00	0.9978
70.00	0.9970
80.00	0.9962
90.00	0.9957
100.00	0.9957
120.00	0.9990
140.00	1.0091
160.00	1.0246

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²