



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/04/29

REF TEK 147A-01/1/H

S/N:81349

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0002
10.00	1.0000
20.00	0.9994
30.00	0.9986
40.00	0.9976
50.00	0.9966
60.00	0.9958
70.00	0.9954
80.00	0.9956
90.00	0.9965
100.00	0.9983
120.00	1.0045
140.00	1.0140
160.00	1.0222

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²