



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/05/14

REF TEK 147A-01/1/H

S/N:81381

P/N: 106293-01

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Horizontal
1.00	1.0000
10.00	1.0000
20.00	0.9999
30.00	0.9997
40.00	0.9996
50.00	0.9995
60.00	0.9997
70.00	1.0000
80.00	1.0007
90.00	1.0017
100.00	1.0033
120.00	1.0089
140.00	1.0207
160.00	1.0392

OUTPUT SENSITIVITY

Horizontal	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²