



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/04/03

REF TEK 147A-01/1/V
S/N:82002

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	1.0000
10.00	1.0000
20.00	1.0000
30.00	1.0000
40.00	0.9999
50.00	0.9997
60.00	0.9991
70.00	0.9985
80.00	0.9978
90.00	0.9973
100.00	0.9973
120.00	1.0007
140.00	1.0122
160.00	1.0305

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²