



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/04/03

REF TEK 147A-01/1/V

S/N:82005

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9992
10.00	1.0000
20.00	1.0021
30.00	1.0049
40.00	1.0074
50.00	1.0087
60.00	1.0082
70.00	1.0059
80.00	1.0018
90.00	0.9972
100.00	0.9928
120.00	0.9888
140.00	0.9979
160.00	1.0199

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²