



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/04/03

REF TEK 147A-01/1/V
S/N:82009

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	1.0006
10.00	1.0000
20.00	0.9984
30.00	0.9960
40.00	0.9931
50.00	0.9902
60.00	0.9877
70.00	0.9861
80.00	0.9857
90.00	0.9871
100.00	0.9906
120.00	1.0048
140.00	1.0287
160.00	1.0522

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²